

# FPFA-T Control Package

**FULLY AUTOMATIC CONTROL PACKAGE**



<b>PURPOSE</b>	To fully automate the use of the <b>FLEXPAK™</b> Leak Detector, allow for recipe storage, repeatability and data acquisition
<b>VACUUM LEVELS</b>	0-27 inches Hg (0-90 kPa) (to atmosphere) 2.92 inches Hg (10 kPa) absolute
<b>ENCLOSURE</b>	Stainless Steel
<b>OPERATOR INTERFACE</b>	4.3 inch (11cm) Touch Screen
<b>CONTROL</b>	Programmable Logic Controller (PLC)
<b>COMPRESSED AIR PRESSURE REQUIRED</b>	70 psi (5 bar) minimum
<b>CONNECTION</b>	¼ NPT; G ¼ adapter supplied outside North America
<b>POWER SUPPLY</b>	120/230 VAC, 50/60 Hz, 1 Phase (power plug provided to suit country of destination)
<b>POWER REQUIRED</b>	40 w
<b>APPROVALS</b>	UL, CSA, CE according to EMC 2014/30EU and Low Voltage Directive 2014/35/EU

## Can be used for the following ASTM test procedures:

**ASTM D3078** Standard Test Method for Determination of Leaks in Flexible Packaging by Bubble Emission

**ASTM D6653** Standard Test Methods for Determining the Effects of High Altitude on Packaging Systems by Vacuum Method

**ASTM D4169** Standard Practice for Performance Testing of Shipping Containers and Systems

**ASTM D4991** Standard Test Method for Leakage Testing of Empty Rigid Containers by Vacuum Method

**ASTM D5094** Standard Test Method for Gross Leakage of Liquids from Containers with Threaded or Lug-Style Closures

**ASTM F2096** Standard Test Method for Gross Leakage of Liquids from Containers with Threaded or Lug-Style Closures

1.416.399.5583

[www.FlexPakInc.com](http://www.FlexPakInc.com) | [sales@FlexPakInc.com](mailto:sales@FlexPakInc.com)

**FLEXPAK™**

FlexPak Leak Detectors Inc.

17769 Centreville Creek Rd. Caledon East, ON L7K 2L9 Canada